

# An Analytical-empirical Calculation of Linear Attenuation Coefficient of Megavoltage Photon Beams

Seif F.<sup>1</sup>, Tahmasebi-Birgani M. J.<sup>2</sup>, Bayatiani M. R.<sup>1\*</sup>

## ABSTRACT

**Background:** In this study, a method for linear attenuation coefficient calculation was introduced.

**Methods:** Linear attenuation coefficient was calculated with a new method that base on the physics of interaction of photon with matter, mathematical calculation and x-ray spectrum consideration. The calculation was done for Cerrobend as a common radiotherapy modifier and Mercury.

**Results:** The values of calculated linear attenuation coefficient with this new method are in acceptable range. Also, the linear attenuation coefficient decreases slightly as the thickness of attenuating filter (Cerrobend or mercury) increased, so the procedure of linear attenuation coefficient variation is in agreement with other documents. The results showed that the attenuation ability of mercury was about 1.44 times more than Cerrobend.

**Conclusion:** The method that was introduced in this study for linear attenuation coefficient calculation is general enough to treat beam modifiers with any shape or material by using the same formalism; however, calculating was made only for mercury and Cerrobend attenuator. On the other hand, it seems that this method is suitable for high energy shields or protector designing.

## Keywords

Cerrobend, Linear Attenuation Coefficient, Energy Spectral

## Introduction

In radiation therapy, beam modifiers such as shield, wedge, compensator and MLC (Multi-Leaf Collimator) are used in order to deliver enough doses to a tumor and protect (Organs at Risk) OARs with achieving desirable dose distribution. In (Intensity Modulated Radiation Therapy) IMRT techniques including MLC base or compensator base, the linear attenuation coefficient is a very important factor to choose benefit beam modifier material and thickness and the quality of IM produced depends on the linear attenuation coefficient [1-6]. In narrow and monoenergy beam conditions, source and detector are assumed to be collimated and the measurement made at a short distance, so the attenuation takes place with just a simple exponential reduction law. In real therapy conditions, polychromatic beams passing through the material with different thicknesses by multiple scattering, it must be entered in treatment planning systems [7-11]. Since X-ray beam of the medical accelerators is broad energy spectrum, it must be considered in attenuation calculations and treatment planning systems [12]. In this study, a new

<sup>1</sup>Assistant professor, Department of Medical Physics and Radiotherapy, Arak University of Medical Sciences, Arak, Iran

<sup>2</sup>Professor, Department of Medical Physics and Radiotherapy, Ahvaz Jundishapur University of Medical Sciences, Ahvaz, Iran

\*Corresponding author: M. R. Bayatiani, Ph.D of Medical Physics, Arak University of Medical Sciences & Radiation Oncology Department of Khansari Hospital, Arak, Iran  
E-mail: m.bayatiani@arakmu.ac.ir

Received: 11 December 2015  
Accepted: 6 March 2016

method was introduced to calculate the linear attenuation coefficient. In this method, linear attenuation coefficient was calculated from integration equation considering the incident energy spectra and changing that happened in beam quality because of various thickness of absorber. Moreover, this calculation will be useful for designing radiation protector materials.

## Material and Methods

### Linear Attenuation Coefficient Calculation

Ideally, linear attenuation coefficient must be measured in narrow beam conditions but in real procedures or broad beams, scattered photons such as Compton photons cannot be ignored. In this paper, linear attenuation coefficient for various thicknesses of Cerrobend and mercury were calculated with the base of experimental dosimetry, mathematical calculations, physics of interaction of photon with matter and with new functional forms for Varian Linac photon spectra [13].

In this research, Cartesian coordinates system has been used because collimator of Varian as like most Linacs are rectangular. As shown in Figure 1, when primary photon beams ( $I_0$ ) passes through the attenuator with thickness  $t_0$  (mercury or Cerrobend in this research) the photon beams that receive the detector ( $I_1$ ) have two components: scattered radiation ( $I_s$ ) and primary radiation ( $I_p$ ). So,

$$I_1 = I_p + I_s \tag{1}$$

When there is not any attenuator filter, the primary photon beams ( $I_0$ ) reduce only because of inverse square law, so in this situation the photon beams that receive the detector ( $I_2$ ) is equal to:

$$I_2 = \frac{I_0}{D^2} \tag{2}$$

Reduction of beams in ( $I_p$ ) is because of attenuator with  $t_0$  thickness and inverse square law, thus ( $I_p$ ) is equal to:

$$I_p = \frac{I_0 e^{-\mu t}}{D^2} \tag{3}$$




Figure 1: schematic diagram of scatter calculation geometry

For calculation of scattered radiation that receives the detector ( $I_s$ ), volume element with  $dx$ ,  $dy$  and  $dz$  dimensions were considered. Since the spectrum of a beam is not monoenergetic, the energy is the other variable. We used 6MV Varian spectrum ( $E$ ) determined with a technique based on the effective spectrum method described in Ref [13].

Thus, the scattered intensity ( $I_s$ ) has been calculated by calculating scatter due to an elemental volume and energy spectrum and integrating over four variables: the thickness of absorber, entire field size and energy spectrum, so we have quadruplet integral. Thus, ( $I_s$ ) is equal to:

$$I_s = \frac{4I_0 \int_0^{t_0} \int_0^{\frac{y_f}{2}} \int_0^{\frac{x_f}{2}} \int_0^{E_c} N_e \frac{e^{-\mu_1}}{r_1} \times \frac{e^{-\mu_2}}{r_2} \left( \frac{d\sigma_0}{d\Omega} F_{KN} \right) \Psi(E) EdEdxdydz}{\int_{E=0}^{E_c} \Psi(E) EdE} \quad (4)$$

In order to simplify equation (4), one can write:

$$I_s = I_0 \text{Int} \quad (5)$$

If the ratio of chamber readings with and without attenuator consider as ( $r$ ), from equations (1 to 5):

$$r = \frac{I_1}{I_2} = \frac{I_0 e^{-\mu} + I_0 \text{Int}}{I_0 / D^2} = e^{-\mu} + D^2 \text{Int} \quad (6)$$

Therefore, with measuring readings ratio for different attenuator thicknesses and solving the integration equation, the linear attenuation coefficient can be calculated for attenuator with arbitrary material or thickness.

### Experimental Dosimetry

The measurement of readings ratio was made on Varian 2100C/D accelerator (Ahvaz Golestan Hospital) at 6 MV photon beams by using absolute dosimetry with Farmer ionization chamber (FC65) and Dose1 electrometer

(scanditronix-wellhöfer) at 110cm distance from source (SSD=110 cm) (Figures 2-4).

Readings ratios were recorded for 6×6, 8×8, 10×10, 12×12 and 14×14 cm<sup>2</sup> field sizes modulated by mercury ( $z=80$ ,  $\rho=13.53$ ) with thicknesses ( $t=0, 0.65, 1, 1.5, 2, 3, 4, 5$  and 6 cm) and Cerrobend ( $z=70.8$ ,  $\rho=9.38$ ) [14] with thicknesses ( $t=0, 1.35, 2.7, 4.05, 5.4, 6.75, 8.1$  and 9.45cm) were placed on the Perspex shielding tray. Chen et al. [15] showed scatter from phantom is significant. In this research, the measurements were performed in air with related build-up cap according to IAEA TRS398 protocol [16]. Since reading ratio is reading of Farmer chamber in the presence of

mercury (or Cerrobend) divided by reading of Farmer chamber without any attenuator at reference dose in water phantom ( $d=10$  cm). The photon energy for all experiments was 6MV. The Cerrobend attenuators have cut with electronic block cutter system that user can have attenuator with desirable thicknesses.

For solving the integration equation, some parameters in equation (4) must be replaced as follow:  $r_0$  is the classical electron radius that is equal to  $2.8182 \times 10^{-13}$  cm,  $\alpha$  is the ratio of photon energy to the electron rest mass energy, so  $\alpha = \frac{h\nu}{m_0 c^2} = \frac{E}{0.511(\text{Mev})}$  that  $E$  is integration

variable and  $N_e$  is the number of electrons/cm<sup>3</sup> that is equal to  $32.49 \times 10^{23}$  for mercury and  $22.92 \times 10^{23}$  for Cerrobend. The other parameter ( $\frac{d\sigma_0}{d\Omega} F_{KN}$ ) is differential of incoherent

scattering cross section per electron and per unit solid angle in direction  $\theta$  that is known as Klein\_Nishina formula and is given by [17].

According to Figure 1 scattering angle,  $\theta$  is equal to:

$$\frac{d\sigma_0}{d\Omega} F_{KN} = \frac{r_0^2}{2} (1 + \cos^2 \theta) F_{KN} = \frac{r_0^2}{2} (1 + \cos^2 \theta) \left\{ \frac{1}{1 + \alpha(1 - \cos \theta)} \right\}^2 \left\{ 1 + \frac{\alpha^2 (1 - \cos \theta)^2}{[1 + \alpha(1 - \cos \theta)](1 + \cos^2 \theta)} \right\} \quad (7)$$

$$\theta = \alpha + \beta \quad (8)$$

$$\theta = \tan^{-1} \frac{D(\sqrt{x^2 + y^2})}{(S - Z)(D - S + Z) - x^2 - y^2} \quad (9)$$

And also  $t_1, t_2, r_1$  and  $r_2$  in Equation 4 must be replaced by:

$$t_1 = \frac{t_0 - z}{S - z} r_1 \text{ where } r_1 = \sqrt{x^2 + y^2 + (D - z)^2} \quad (10)$$

$$t_2 = \frac{z}{D - S + z} r_2 \text{ where } r_2 = \sqrt{x^2 + y^2 + (D - S + z)^2} \quad (11)$$

As mentioned before, a typical therapy beam is not monoenergetic, we replaced ( $E$ ) in integration equation with Varian Linac 6MV photon spectra functional form that is specified by Ali et al. [16].

$$\psi(E) = \left[ 1 + C_3 + \left( \frac{E}{E_0} \right)^2 \right] \left[ \ln \left( \frac{E_e (E_e - E)}{E} + 1.65 \right) - 0.5 \right] \exp(-\mu_w(E)C_1^2 - \mu_{Al}(E)C_2^2) \quad (12)$$

For Varian 6MV photons  $C_1=1.222$ ,  $C_2=5.147$ ,  $C_3=-1.186$  and  $E_e=5.76$  Mev

Parameterization of  $\mu_w$  and  $\mu_{Al}$  is extracted from table 3 of Ali's article. At the end, with replace all necessary parameters and equations and readings ratio from dosimetry in Equation (6); we have four variables (three for coordinates and one for energy) in integration equation. This is a complicated integration equation because attenuation coefficient is related to the energy and also is a part of integrand, so there is not simple or direct method to solve this integrated equation consisting of quadruplet integral. We solved this by Gaussian integration method in Matlab software and calculated attenuation coefficients for different thicknesses of mercury and Cerrobend.

## Results

### Linear Attenuation Coefficient Calculation

Linear attenuation coefficient of variable attenuator and field sizes calculated for 6MV photon beam from new method by solving a complicated integration equation consist of quadruplet integral described before. The values of Reading ratio and calculated linear attenuation coefficient for different Cerrobend and mercury thickness were tabulated in Tables 1 and 2.

The linear attenuation coefficient values of different Cerrobend and mercury thicknesses that were calculated for 6×6, 8×8, 10×10, 12×12 and 14×14 cm<sup>2</sup> field sizes are shown in Tables 1 and 2.



Figure 2: Farmer Ionization chamber (FC65)

**Table 1:** Reading ratio and calculated linear attenuation coefficient for different Cerrobend thickness

Cerrobend thickness (cm±0.01)	6×6 cm <sup>2</sup>		8×8 cm <sup>2</sup>		10×10 cm <sup>2</sup>		12×12 cm <sup>2</sup>		14×14 cm <sup>2</sup>	
	r	mu	r	mu	R	mu	r	mu	r	mu
1.35	48.5	0.514	48.8	0.491	49	0.472	49.3	0.447	51.2	0.395
2.7	26	0.477	26.3	0.457	27	0.427	26.9	0.405	28	0.364
4.05	14.6	0.453	14.8	0.433	15	0.409	15.3	0.378	16	0.337
5.4	8.3	0.439	8.4	0.419	8.6	0.392	8.8	0.359	9.2	0.315
6.75	4.7	0.431	4.8	0.409	4.9	0.382	5	0.346	5.3	0.293
8.1	2.7	0.423	2.8	0.4	2.9	0.369	2.9	0.331	3.1	0.28
9.45	1.3	0.42	1.4	0.4	1.4	0.368	1.5	0.33	1.6	0.27

**Table 2:** Reading ratio and calculated linear attenuation coefficient for different Mercury thickness

Mercury thickness (cm±0.01)	6×6 cm <sup>2</sup>		8×8 cm <sup>2</sup>		10×10 cm <sup>2</sup>		12×12 cm <sup>2</sup>		14×14 cm <sup>2</sup>	
	R	mu	r	mu	R	mu	R	mu	r	Mu
0.65	58.6	0.792	59	0.759	59	0.733	58.8	0.709	60.9	0.623
1	46.9	0.726	47	0.702	47.5	0.665	47.8	0.628	49.5	0.56
1.5	34	0.688	34.6	0.654	35	0.619	35.1	0.585	36.4	0.525
2	24.7	0.668	25	0.639	25.3	0.605	25.5	0.567	26.6	0.508
3	13.6	0.634	13.8	0.605	14	0.569	14.2	0.527	14.9	0.467
4	7	0.633	7.5	0.591	7.7	0.551	7.9	0.503	8.3	0.437
5	4	0.612	4.3	0.571	4.4	0.531	4.5	0.478	4.8	0.397
6	2	0.619	2.4	0.562	2.5	0.516	2.6	0.452	2.7	0.38



**Figure 3:** Dose 1 Electrometer

### Discussion

When a beam modifier (compensator or shield) is placed in an x-ray beam, the characteristics of the beam change because of scattered photon produced in the beam modifier. This work studied the influence of scattered photons from Cerrobend and mercury modifiers on the transmission and the quality changes in beam. In addition, we introduced a method for linear attenuation coefficient calculations.

The exponential model is usually used for modifiers attenuation calculation [18]; however, this model introduces unnecessary errors




**Figure 4:** Varian Linac and water phantom

for polyenergetic beams. On the other hand, the linear attenuation coefficient is strongly energy dependent. In general, lower energetic X-ray photons have a higher interaction probability. Since an X-ray device produces photons in a wide energy range, the transmission should actually be considered for the whole energy range. Alles et al. [12] suggested the model for linear attenuation coefficient calculations for different discretizing thicknesses and energies (up to 150 keV). The method that was introduced in this paper was based on Equation 6 considered all energies and thickness continuously. The procedure of calculated linear attenuation coefficient from method of this study versus attenuator thickness is in agreement with other studies such as du Ples-


sis et al [3] and Sasaki et al. [10]. As shown in Figure 5 and 6, the linear attenuation coefficient decreases slightly as the thickness of the attenuating filter (Cerrobend or mercury) increases, reflecting the progressive hardening of X-ray beam. These data indicate that linear attenuation coefficients are approximately linear with modifier thickness increment up to about 4cm for cerrobend and 2.8 cm for mercury (this reported about 3cm for MCP96 alloy) [8] which shows hardening effect happened at lower energy but for much thicker modifiers, the change in linear attenuation coefficient did not occur. It is also seen that the value of linear attenuation coefficient decreases as the field size increases, because the scatter component increases at large field sizes and can reach the detector. It is seen that the curves are parallel to each other. On the other hand, the difference between the linear attenuation coefficient for  $6 \times 6 \text{ cm}^2$  and  $14 \times 14 \text{ cm}^2$  is approximately constant for all Cerrobend and mercury thicknesses.

## Conclusion

The method introduced in this study for linear attenuation coefficient calculation is general enough to treat compensatory filters, wedges, shields and any intensity modulator or radiation absorber with any shape or ma-



**Figure 5:** Linear attenuation coefficient values v.s mercury attenuator thicknesses for different field sizes.



**Figure 6:** linear attenuation coefficient for different thickness of mercury attenuator and field sizes.

terial by using the same formalism, although calculating was made only for mercury and Cerrobend attenuator. On the other hand, it seems that this method is suitable for high energy shields or protector designing.

### Acknowledgment

This work was supported by Ahvaz Jundishapur University of Medical Sciences [Grant number: u-90083].

### Conflict of Interest

None

### References

1. Chang SX, Cullip TJ, Deschesne KM. Intensity modulation delivery techniques: "step & shoot" MLC auto-sequence versus the use of a modulator. *Med Phys*. 2000;**27**:948-59. doi.org/10.1118/1.598989. PubMed PMID: 10841397.
2. Meyer J, Mills JA, Haas OC, Parvin EM, Burnham KJ. Some limitations in the practical delivery of intensity modulated radiation therapy. *Br J Radiol*. 2000;**73**:854-63. doi.org/10.1259/bjr.73.872.11026861. PubMed PMID: 11026861.
3. du Plessis FC, Willemse CA. Monte Carlo calculation of effective attenuation coefficients for various compensator materials. *Med Phys*. 2003;**30**:2537-44. doi.org/10.1118/1.1591432. PubMed PMID: 14528976.
4. Midgley SM. Materials analysis using x-ray linear attenuation coefficient measurements at four photon energies. *Phys Med Biol*. 2005;**50**:4139-57. doi.org/10.1088/0031-9155/50/17/016. PubMed PMID: 16177536.
5. Lin JP, Chu TC, Liu MT. Dose compensation of the total body irradiation therapy. *Appl Radiat Isot*. 2001;**55**:623-30. doi.org/10.1016/S0969-8043(01)00129-4. PubMed PMID: 11573795.
6. Tahmasebi-Birgani MJ, Seif F, Bayatiani MR. Dosimetric characteristics of mercury and cerrobend blocks in megavoltage radiation therapy. *Journal of Radioanalytical and Nuclear Chemistry*. 2015;**303**:1843-50.
7. Huang PH, Chin LM, Bjarngard BE. Scattered photons produced by beam-modifying filters. *Med Phys*. 1986;**13**:57-63. doi.org/10.1118/1.595923. PubMed PMID: 3951410.
8. Mejaddem Y, Hyodynmaa S, Brahme A. Photon scatter in intensity modulating filters evaluated by first Compton scatter and Monte Carlo calculations and experiments in broad beams. *Phys Med Biol*. 2000;**45**:2747-60. doi.org/10.1088/0031-9155/45/10/302. PubMed PMID: 11049169.
9. Dimitriadis D, Fallone B. Compensators for intensity-modulated beams. *Medical Dosimetry*. 2002;**27**:215-20. doi.org/10.1016/S0958-3947(02)00139-5. PubMed PMID: 17592450.
10. Sasaki K, Obata Y. Dosimetric characteristics of a cubic-block-piled compensator for intensity-modulated radiation therapy in the Pinnacle radiotherapy treatment planning system. *J Appl Clin Med Phys*. 2007;**8**:85-100. PubMed PMID: 17592450.
11. Midgley SM. A parameterization scheme for the x-ray linear attenuation coefficient and energy absorption coefficient. *Phys Med Biol*. 2004;**49**:307-

25. doi.org/10.1088/0031-9155/49/2/009. PubMed PMID: 15083673.
12. Alles J, Mudde RF. Beam hardening: analytical considerations of the effective attenuation coefficient of X-ray tomography. *Med Phys*. 2007;**34**:2882-9. doi.org/10.1118/1.2742501. PubMed PMID: 17821996.
13. Ali ES, Rogers DW. Functional forms for photon spectra of clinical linacs. *Phys Med Biol*. 2012;**57**:31-50. doi.org/10.1088/0031-9155/57/1/31. PubMed PMID: 22126713.
14. Birgani MJT, Seif F, Chegeni N, Bayatiani MR. Determination of the effective atomic and mass numbers for mixture and compound materials in high energy photon interactions. *Journal of Radioanalytical and Nuclear Chemistry*. 2012;**292**:1367-70. doi.org/10.1007/s10967-012-1677-2.
15. Chen Z, Wang X, Bortfeld T, Mohan R, Reinstein L. The influence of scatter on the design of optimized intensity modulations. *Med Phys*. 1995;**22**:1727-33. doi.org/10.1118/1.597536. PubMed PMID: 8587525.
16. IAEA. Absorbed Dose Determination in Electron Beam Radiotherapy: An International Code of Practice for Dosimetry Based on Standards of Absorbed Dose to Water IAEA Technical Report Series, IAEA TRS-398; 2001.
17. Johns HE. Physics of radiology: Charles River Media; 1983.
18. Gürler O, Yalçın S. A practical method for calculation of mass-attenuation coefficients of  $\beta$  particles. *Annals of Nuclear Energy*. 2005;**32**:1918-25. doi.org/10.1016/j.anucene.2005.05.007.